

Journal of Applied Oral Science (2022) 30:e2022er002 – ISSN 1678-7765 http://dx.doi.org/10.1590/1678-7757-2022er002

ERRATUM

Due to a publishing error the article: "The effect of cooling procedures on monomer elution from heatcured polymethyl methacrylate denture base materials", published at Journal of Applied Oral Science 2022;30:e20220161 was printed with the following error:

Where it reads:

Material/lot	Code	Composition/curing method*	Manufacturer	
Paladon 65		Powder: Methacrylate copolymonomers (0-5%), BPO <1%		
Powder: 010164		Liquid: MMA (>90%), BDMA (0-5%)	Heraeus Kulzer GmbH, Hanau,	
Liquid: R010048	PA	Powder/Liquid ratio: 10 g/4 mL	Germany	
]	Curing: Water, 80°C/15 min + 100°C/20 min		
ProBase Hot		Powder: PMMA (>95%), BPO (1-1.5%)		
Powder: UN0447		Liquid: MMA (50-100%), EGDMA (3-10%)	Ivoclar-Vivadent, Schaan,	
Liquid: V33082		Powder/Liquid ratio: 22.5 g/10 mL	Liechtenstein	
		Curing: Boiling water for 45 min		
Stellon QC-20		Powder: PMMA, BPO		
Powder:130CT053		Liquid: MMA, HQ (0.01%)	DeguDent GmbH, Hanau-	
Liquid: 13JUL117	QC	Powder/Liquid ratio: 24 g/10 mL	Wolfgang, Germany	
		Curing: Water, 100oC/20 min		
Vertex Rapid Simplified		Powder: PMMA (>99%), accelerator (<1%), coloring agents (<1%)		
Powder: XT382P02	VE	Liquid: MMA (>95%), crosslinker (<5%), accelerator (<1%)	Vertex Dental B.V, Zeist,	
Liquid: XT381L02] _	Powder/Liquid ratio: 2.3 g/1 mL	I he Netherlands	
		Curing: Water, 100oC/20 min		

* According to the manufacturers' information. MMA: Methyl methacrylate, PMMA: Polymethyl methacrylate, BPO: Benzoyl peroxide, HQ: Hydroquinone, EGDMA: Ethyleneglycol dimethacrylate, BDMA: Tetramethylene dimethacrylate

Figure 1- The heat-cured denture base materials used in the study

It should read:

Material/lot	Code	Composition/curing method*	Manufacturer	
Paladon 65		Powder: Methacrylate copolymonomers (0-5%), BPO <1%		
Powder: 010164		Liquid: MMA (>90%), BDMA (0-5%)	Heraeus Kulzer GmbH, Hanau,	
Liquid: R010048	PA	Powder/Liquid ratio: 10 g/4 mL	Germany	
		Curing: Water, 80°C/15 min + 100°C/20 min		
ProBase Hot		Powder: PMMA (>95%), BPO (1-1.5%)		
Powder: UN0447		Liquid: MMA (50-100%), EGDMA (3-10%)	Ivoclar-Vivadent, Schaan,	
Liquid: V33082	РВ	Powder/Liquid ratio: 22.5 g/10 mL	Liechtenstein	
		Curing: Boiling water for 45 min		
Stellon QC-20		Powder: PMMA, BPO		
Powder:130CT053	00	Liquid: MMA, HQ (0.01%)	DequDent GmbH, Hanau-	
Liquid: 13JUL117	QC	Powder/Liquid ratio: 24 g/10 mL	Wolfgang, Germany	
		Curing: Water, 100oC/20 min		
Vertex Rapid Simplified		Powder: PMMA (>99%), accelerator (<1%), coloring agents (<1%)		
Powder: XT382P02	VE	Liquid: MMA (>95%), crosslinker (<5%), accelerator (<1%)	Vertex Dental B.V, Zeist,	
Liquid: XT381L02		Powder/Liquid ratio: 2.3 g/1 mL		
		Curing: Water, 100oC/20 min		

* According to the manufacturers' information. MMA: Methyl methacrylate, PMMA: Polymethyl methacrylate, BPO: Benzoyl peroxide, HQ: Hydroquinone, EGDMA: Ethyleneglycol dimethacrylate, BDMA: Tetramethylene dimethacrylate

Figure 1- The heat-cured denture base materials used in the study

(cc) BY

This is an Open Access article distributed under the terms of the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

Journal of Applied Oral Science (2022) 30:e2022er002 – ISSN 1678-7765 http://dx.doi.org/10.1590/1678-7757-2022er002

Where it reads:

Cooling procedures	MMA eluted in water (ppm)							
	PA	PB	QC	VE				
А	4.6 <loq< td=""><td>4.6<loq< td=""><td>8.5 (1.7)^{a,A}</td><td>6.4 (0.3)^{b,A}</td></loq<></td></loq<>	4.6 <loq< td=""><td>8.5 (1.7)^{a,A}</td><td>6.4 (0.3)^{b,A}</td></loq<>	8.5 (1.7) ^{a,A}	6.4 (0.3) ^{b,A}				
В	<lod< td=""><td>2.1<loq< td=""><td><lod< td=""><td>8.7 (2.4)^b</td></lod<></td></loq<></td></lod<>	2.1 <loq< td=""><td><lod< td=""><td>8.7 (2.4)^b</td></lod<></td></loq<>	<lod< td=""><td>8.7 (2.4)^b</td></lod<>	8.7 (2.4) ^b				
С	<lod< td=""><td>4.8<loq< td=""><td>13.2 (2.4)ª</td><td>2.9<loq< td=""></loq<></td></loq<></td></lod<>	4.8 <loq< td=""><td>13.2 (2.4)ª</td><td>2.9<loq< td=""></loq<></td></loq<>	13.2 (2.4)ª	2.9 <loq< td=""></loq<>				
D	2.8 <loq< td=""><td>2.2<loq< td=""><td><lod< td=""><td>3.2<loq< td=""></loq<></td></lod<></td></loq<></td></loq<>	2.2 <loq< td=""><td><lod< td=""><td>3.2<loq< td=""></loq<></td></lod<></td></loq<>	<lod< td=""><td>3.2<loq< td=""></loq<></td></lod<>	3.2 <loq< td=""></loq<>				
E	<lod< td=""><td><lod< td=""><td><lod< td=""><td>4.2<loq< td=""></loq<></td></lod<></td></lod<></td></lod<>	<lod< td=""><td><lod< td=""><td>4.2<loq< td=""></loq<></td></lod<></td></lod<>	<lod< td=""><td>4.2<loq< td=""></loq<></td></lod<>	4.2 <loq< td=""></loq<>				
Control	<lod< td=""><td><lod< td=""><td><lod< td=""><td><lod< td=""></lod<></td></lod<></td></lod<></td></lod<>	<lod< td=""><td><lod< td=""><td><lod< td=""></lod<></td></lod<></td></lod<>	<lod< td=""><td><lod< td=""></lod<></td></lod<>	<lod< td=""></lod<>				

 Table 1- The results of the MMA concentration in the watereluents of the heat-cured PMMA denture base materials tested*

* Means and standand deviations (in parentheses).Superscript letters show mean values with insignificant differences within each material group (lower case) and between material groups per treatment (upper case). LoQ: Lower limit of quantitation (5.90 ppm), LoD: Limit of detection (1.95 ppm). Bold characters show the values obtained using the cooling modes suggested by the manufacturers. Data given for results <LoQ represent only mean values

It should read:

Cooling procedures	MMA eluted in water (ppm)						
	PA	PB	QC	VE			
А	4.6 <loq< td=""><td>4.6<loq< td=""><td>8.5 (1.7)^{a,A}</td><td>6.4 (0.3)^{b,A}</td></loq<></td></loq<>	4.6 <loq< td=""><td>8.5 (1.7)^{a,A}</td><td>6.4 (0.3)^{b,A}</td></loq<>	8.5 (1.7) ^{a,A}	6.4 (0.3) ^{b,A}			
В	<lod< td=""><td>2.1<loq< td=""><td><lod< td=""><td>8.7 (2.4)^b</td></lod<></td></loq<></td></lod<>	2.1 <loq< td=""><td><lod< td=""><td>8.7 (2.4)^b</td></lod<></td></loq<>	<lod< td=""><td>8.7 (2.4)^b</td></lod<>	8.7 (2.4) ^b			
С	<lod< td=""><td>4.8<loq< td=""><td>13.2 (2.4)^a</td><td>2.9<loq< td=""></loq<></td></loq<></td></lod<>	4.8 <loq< td=""><td>13.2 (2.4)^a</td><td>2.9<loq< td=""></loq<></td></loq<>	13.2 (2.4) ^a	2.9 <loq< td=""></loq<>			
D	2.8 <loq< td=""><td>2.2<loq< td=""><td><lod< td=""><td>3.2<loq< td=""></loq<></td></lod<></td></loq<></td></loq<>	2.2 <loq< td=""><td><lod< td=""><td>3.2<loq< td=""></loq<></td></lod<></td></loq<>	<lod< td=""><td>3.2<loq< td=""></loq<></td></lod<>	3.2 <loq< td=""></loq<>			
E	<lod< td=""><td><lod< td=""><td><lod< td=""><td>4.2<loq< td=""></loq<></td></lod<></td></lod<></td></lod<>	<lod< td=""><td><lod< td=""><td>4.2<loq< td=""></loq<></td></lod<></td></lod<>	<lod< td=""><td>4.2<loq< td=""></loq<></td></lod<>	4.2 <loq< td=""></loq<>			
Control	<lod< td=""><td><lod< td=""><td><lod< td=""><td><lod< td=""></lod<></td></lod<></td></lod<></td></lod<>	<lod< td=""><td><lod< td=""><td><lod< td=""></lod<></td></lod<></td></lod<>	<lod< td=""><td><lod< td=""></lod<></td></lod<>	<lod< td=""></lod<>			

 Table 1- The results of the MMA concentration in the watereluents of the heat-cured PMMA denture base materials tested*

* Means and standand deviations (in parentheses).Superscript letters show mean values with insignificant differences within each material group (lower case) and between material groups per treatment (upper case). LoQ: Lower limit of quantitation (5.90 ppm), LoD: Limit of detection (1.95 ppm). Bold characters show the values obtained using the cooling modes suggested by the manufacturers. Data given for results <LoQ represent only mean values